



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

1 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	1&3 Phase Energy (240V, 5A & UPF)@ 50Hz	Using Accucheck and energy source by Direct/ comparison Method	1.2 Wh to 3.6 kWh	1.18%
2	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (10Hz-1kHz)	Using 8 ½ DMM FLUKE and MFC by Direct/ Comparison Method	10 A to 20 A	0.15 % to 0.12 %
3	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (10Hz-1kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	100 µA to 1 A	0.55 % to 0.20 %
4	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (10Hz-5kHz)	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	1 mA to 200 mA	0.09 % to 0.06 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

2 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (10Hz-5kHz)	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 µA to 200 µA	0.3 % to 0.05 %
6	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (10Hz-5kHz)	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	200 µA to 1 mA	0.05 % to 0.09 %
7	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (10Hz-5kHz)	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	200 mA to 10 A	0.06 % to 0.15 %
8	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (1kHz-5kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	1 A to 3 A	0.20 % to 0.25 %
9	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (1kHz-5kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	100 µA to 1 A	0.55 % to 0.20 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

3 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
10	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC High Current @ (50Hz)	Using C.T.+6 ½ DMM FLUKE & current source by Direct/ comparison Method	20 A to 2000 A	2.10 % to 2.40 %
11	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC High Voltage @ (50Hz)	Using HV Divider & HV source by Direct/ comparison Method	1 kV to 20 kV	2.40 % to 2.60 %
12	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage @ (10Hz-10kHz)	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	1 V to 1000 V	0.041 % to 0.02 %
13	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage @ (10Hz-10kHz)	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 mV to 200 mV	0.20 % to 0.025 %
14	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage @ (10Hz-10kHz)	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	200 mV to 1 V	0.025 % to 0.041 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

4 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
15	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage @ (10Hz-20kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 mV to 1000 V	0.90 % to 0.15 %
16	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage @ (10kHz-100kHz)	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 mV to 20 V	0.32 % to 0.08 %
17	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage @ (10kHz-100kHz)	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	20 V to 100 V	0.08 % to 0.15 %
18	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage @ (20kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 mV to 700 V	0.75 % to 0.25 %
19	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Capacitance @ (1kHz)	Using 6 ½ DMM FLUKE & capacitance box by Direct/ Comparison Method	10 mF to 100 mF	1.85 % to 4.90 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

5 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
20	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Capacitance @ (1kHz)	Using 6 ½ DMM FLUKE & capacitance box by Direct/ Comparison Method	1nF to 10 mF	5.47 % to 1.85 %
21	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current @ (10Hz-1kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	30 µA to 3 A	0.62 % to 0.09 %
22	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current @ (1kHz-5kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	30 µA to 330 mA	0.95 % to 0.2 %
23	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current @ (1kHz-5kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 mA to 10 A	0.2 % to 3.5 %
24	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current @ (45Hz-1kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	3 A to 20 A	0.09 % to 0.21 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

6 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
25	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current @ (50Hz-60Hz)	Using 5522A/9100 Calibrator(Fluke) with CC by Direct Method	10 A to 1000 A	0.5 % to 0.55 %
26	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Power 50Hz-60Hz/ 0.2pF to 1pF(Lead & Lag)/ 30V to 1000V/ 10mA to 20A	Using 5522A calibrator(Fluke)/3050 calibrator by Direct method	60 mW to 20 kW	1.8 % to 0.38 %
27	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage @ (10Hz-45Hz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	1 mV to 33 V	0.9 % to 0.05 %
28	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage @ (10kHz-100kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	30 mV to 330 mV	0.5 % to 0.15 %
29	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage @ (10kHz-100kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 mV to 330 V	0.15 % to 0.32 %
30	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage @ (45Hz-10kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	1 mV to 330 mV	0.8 % to 0.02 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

7 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
31	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage @ (45Hz-10kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 mV to 1000 V	0.02 % to 0.04 %
32	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance @ (1kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	220 pF to 330 nF	5.85 % to 0.45 %
33	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance @ (1kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 nF to 30 mF	0.45 % to 1.5 %
34	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Inductance @ (1kHz)	Using Inductance box by Direct Method	1 mH to 10 H	3%
35	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 µA to 20 mA	0.052 % to 0.003 %
36	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using Shunt with 6 ½ DMM FLUKE & current source by Direct/ Comparison Method	10 A to 100 A	0.9 % to 0.53 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

8 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
37	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	100 µA to 100 mA	0.1 % to 0.07 %
38	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	100 mA to 10 A	0.07 % to 0.2 %
39	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	20 mA to 20 A	0.003 % to 0.05 %
40	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC High Voltage	Using HV Divider with indicator &HV source by Direct/ comparison Method	1 kV to 20 kV	1.97 % to 2.10 %
41	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	1 mV to 1 V	0.42 % to 0.085 %
42	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	1 V to 1000 V	0.085 % to 0.006 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

9 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
43	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 µV to 1000 V	1.75 % to 0.0008 %
44	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	Using 8 ½ DMM FLUKE & resistance box by Direct/ Comparison Method	1 mohm to 2 Mohm	0.5 % to 0.0015 %
45	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	Using 6 ½ DMM FLUKE & resistance box/ MFC by Direct/ Comparison Method	1 ohm to 100 ohm	0.15 % to 0.07 %
46	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	Using 6 ½ DMM FLUKE, & resistance box / MFC by Direct/ Comparison Method	100 ohm to 1 Gohm	0.07 % to 2.6 %
47	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	Using 8 ½ DMM FLUKE & resistance box / MFC by Direct/ Comparison Method	2 Mohm to 20 Mohm	0.01 % to 0.0035 %
48	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	Using 8 ½ DMM FLUKE & resistance box by Direct/ Comparison Method	20 Mohm to 20 Gohm	0.0035 % to 0.3 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

10 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
49	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using 5522A/9100 Calibrator(Fluke) by Direct Method	10 μ A to 330 mA	0.25 % to 0.02 %
50	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using 5522A/9100 Calibrator(Fluke) by Direct Method	10 A to 20 A	0.07 % to 0.026 %
51	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using 5522A/9100 Calibrator(Fluke) with CC by Direct Method	20 A to 1000 A	0.45 % to 0.64 %
52	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 mA to 10 A	0.02 % to 0.07 %
53	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using 5522A/9100 Calibrator(Fluke) by Direct Method	1 mV to 33 V	0.15 % to 0.002 %
54	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using 5522A/9100 Calibrator(Fluke) by Direct Method	33 V to 1000 V	0.002 % to 0.003 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

11 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
55	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	Using 5522A/9100 Calibrator(Fluke) by Direct Method	1 Mohm to 10 Mohm	0.004 % to 0.03 %
56	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	Using Standard Resistance Box by Direct Method	1 Gohm to 200 Gohm	1.74 % to 3.5 %
57	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	Using 5522A/9100 Calibrator(Fluke) by Direct Method	1 ohm to 1 Mohm	0.1 % to 0.004 %
58	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	Using 5522A/9100 Calibrator(Fluke) by Direct Method	10 Mohm to 330 Mohm	0.03 % to 0.35 %
59	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 Mohm to 1 Gohm	0.35 % to 0.2 %
60	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscopes/ Amplitude	Using 5522A (Fluke)/3050 Calibrator by Direct Method	1 mV to 130 V	4.9 % to 0.35 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

12 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
61	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscopes/Bandwidth	Using 5522A (Fluke)/3050 Calibrator by Direct Method	50 kHz to 1.1 GHz	2.5%
62	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscopes/Time	Using 5522A (Fluke)/3050 Calibrator by Direct Method	2 ns to 5 s	0.03 % to 0.6 %
63	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	B-Type Thermocouple	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	100 °C to 1800 °C	0.6°C
64	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	E-Type Thermocouple	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	-200 °C to 1000 °C	0.087°C
65	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	J-Type Thermocouple	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	-200 °C to 1000 °C	0.08°C
66	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	K-Type Thermocouple	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method.	-200 °C to 1200 °C	0.085°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

13 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
67	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	N-Type Thermocouple	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	-200 °C to 1300 °C	0.13°C
68	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	R&S-Type Thermocouple	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	0 to 1700 °C	0.6°C
69	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	RTD	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	-200 °C to 800 °C	0.253°C
70	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	T-Type Thermocouple	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	-200 °C to 400 °C	0.14°C
71	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	B-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	450 °C to 1820 °C	0.8°C
72	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	E-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 1000 °C	0.15°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

14 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
73	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	J-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 1000 °C	0.6°C
74	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	K-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 1200 °C	0.6°C
75	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	N-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 1300 °C	0.6°C
76	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	R&S-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	100 °C to 1700 °C	0.65°C
77	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	RTD	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 800 °C	0.085°C
78	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	T-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 400 °C	0.21°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2412 **Page No** 15 of 53

Validity 10/09/2020 to 09/09/2022 **Last Amended on** -

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
79	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Function Generator & MFC by Direct/ comparison Method	1 MHz to 10 MHz	0.006 % to 0.06 %
80	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using 8 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 Hz to 1 MHz	0.06 % to 0.006 %
81	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 Hz to 1000 kHz	0.06%
82	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Timer/Stop Watch/Hour meter/Time totalizer	Using time totalizer by comparison Method	2 s to 30 min	0.35 s to 1 s
83	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Timer/Stop Watch/Hour meter/Time totalizer	Using time totalizer by comparison Method	30 min to 24 hr	1 s to 56 s
84	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Frequency	Using 5522A/9100 Calibrator (Fluke)/Function Generator by Direct Method	10 Hz to 10 MHz	0.001 % to 0.15 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

16 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
85	MECHANICAL-ACCELERATION AND SPEED	Speed -RPM Tachometer (Contact)	Using Tachometer (source RPM generator/Calibrator)by comparison Method	>1000 RPM to 4000 RPM	10RPM
86	MECHANICAL-ACCELERATION AND SPEED	Speed -RPM Tachometer (Contact)	Using Tachometer (source RPM generator/Calibrator)by comparison Method	100 RPM to 1000 RPM	3.5RPM
87	MECHANICAL-ACCELERATION AND SPEED	Speed -RPM Tachometer (Non-Contact) / Stirrer	Using Tachometer (source RPM generator/Calibrator)by comparison Method	>4000 RPM to 47000 RPM	18RPM
88	MECHANICAL-ACCELERATION AND SPEED	Speed -RPM Tachometer (Non-Contact) / Stirrer	Using Tachometer (source RPM generator/Calibrator)by comparison Method	100 RPM to 4000 RPM	3.7RPM
89	MECHANICAL-ACOUSTICS	Sound Level Meter(1kHz)	Sound level generator calibrator Direct method	94dB and to 114 dB	1.8dB
90	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bevel Protector L.C.- 5 min	Using Angle Slip Gauge as per IS 5812	0-90-0 °	6.5min of arc



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

17 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
91	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bore Gauge with or without Dial - Transmission stroke (All ranges) L.C-0.001 mm	Using Dial calibration tester by comparison method	0 to 2 mm	3µm
92	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Coating Thickness Gauge L.C. 1 µm	Using Coating thickness foils by comparison method	0 to 1.20 mm	10µm
93	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Combination Set L.C.-1 °	Using Angle Slip Gauge as per IS 5812	0-90-0 °	35 min of arc
94	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Micrometer L.C.- 0.001mm	Using Slip gauge & Accessories by comparison method	0 to 300 mm	7µm
95	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Vernier Caliper L.C.-0.01 mm	Using Slip gauge & Accessories	0 to 300 mm	15µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC.,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

18 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
96	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Vernier Caliper L.C.-0.02 mm	Using Slip gauge & Accessories	0 to 600 mm	25µm
97	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Calibration Tester L.C.0.1 µm	Using Sylvac probe with DRO by comparison Method	0 to 25 mm	1.5µm
98	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial/Digital Thickness Gauge L.C-0.001 mm	Using Slip gauge & Accessories by comparison method	0 to 1 mm	2µm
99	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial/Digital Thickness Gauge L.C-0.01 mm	Using Slip gauge & Accessories by comparison	0 to 25 mm	7.60µm
100	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Electronic Probe With DRO / Comparator L.C.-0.0001 mm	Using Slip gauge set by Comparison method	0 to 25 mm	1µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC.,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

19 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
101	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer L.C.-0.001mm	Using Slip gauge	0 to 100 mm	2µm
102	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer L.C.-0.01mm	Using Slip gauge & Accessories	> 100 mm to 300 mm	8µm
103	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer L.C.-0.01mm	Using Slip gauge & Accessories	> 300 mm to 1000 mm	15µm
104	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Feeler Gauge	Using Sylvac probe with DRO by Comparison method	0.01 mm to 1 mm	2.50µm
105	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Height Gauge/Micro Height Gauge (Vernier/Dial/Digital) L.C.-0.1 µm	Using Slip gauge	0 to 1000 mm	10µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

20 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
106	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer L.C.-0.01 mm	Using Slip gauge & DRO with probe	> 300 mm to 1000 mm	13µm
107	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer L.C.-0.01 mm	Using Slip gauge & DRO with probe	0.01 mm to 300 mm	8µm
108	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Lever Dial L.C.-0.001 mm	Using Dial calibration tester	0 to 2 mm	2µm
109	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Pin Gauge	Using Sylvac probe with Comparator stand	0.10 mm to 20 mm	2µm
110	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Scale L.C.-0.5 mm	Using Tape & Scale Calibrator by comparison method.	0 to 2000 mm	118v Lµm,where L in mm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

21 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
111	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Tape/Pie Tape L.C.-1 mm	Using Tape & Scale Calibrator by comparison method.	0 to 100 m	118vLµm,where L in mm
112	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Head/Drum L.C.0.001 mm	Using Sylvac probe with DRO as per IS 9483	0 to 25 mm	1.50µm
113	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Rod/Height Master	Using Slip gauge & sylvac probe with DRO by comparison method	> 275 mm to 1000 mm	9.15µm
114	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Rod/Height Master	Using Slip gauge & sylvac probe with DRO by comparison method	Up to 275 mm	3.50µm
115	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Paint Thickness Gauge Foils	Using Sylvac probe with DRO by comparison method	0.005 mm to 10 mm	1.50µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

22 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
116	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Pistol Caliper L.C-0.01 mm	Using Slip gauge by comparison method	0 to 150 mm	60µm
117	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Ring Gauge	Using LMM 300 & Master Ring Gauge	> 100 mm to 300 mm	4µm
118	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Ring Gauge	Using LMM 300 & Master Ring Gauge	3 mm to 100 mm	2µm
119	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain/Paddle Plug/Master/Keyway Gauges	Using Slip gauge& sylvac probe with DRO	> 100 mm to 300 mm	4µm
120	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain/Paddle Plug/Master/Keyway Gauges	Using Slip gauge& sylvac probe with DRO	0.50 mm to 100 mm	2.30µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

23 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
121	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Dial L.C.-0.001 mm	Using Dial calibration tester	0 to 25 mm	3µm
122	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Dial L.C.-0.001 mm	Using Slip gauge	0 to 50 mm	6µm
123	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Precision Spirit Level Sensitivity 0.02 mm/m Base Length 300 mm	Using Electronic Level by comparison method	0 to +/-1 mm/m	25µm/m
124	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Snap Gauge/Gap Gauge	Using Slip gauge	0.50 mm to 100 mm	3µm
125	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Snap Gauge/Gap Gauge	Using LMM 300	100 mm to 300 mm	5µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

24 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
126	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Squareness of Tri Square/Engineering Square	Using Master square cylinder ,Slip Gauges by Comparison Method	Up to 600 mm	15µm
127	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Plug Gauge (Effective ,Minor,Major Dia)	Using LMM 300	> 100 mm to 300 mm	4µm
128	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Plug Gauge (Effective ,Minor,Major Dia)	Using FCDM m/c	2 mm to 100 mm	4µm
129	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Ring Gauge (Effective Dia)	Using LMM 300 & Master Ring Gauge	> 100 mm to 300 mm	5µm
130	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Ring Gauge (Effective Dia)	Using LMM 300 & Master Ring Gauge	3 mm to 100 mm	2µm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

25 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
131	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	V-Block-Symmetry, Parallelism & Squareness	Using Surface Plate, Test Mandrels & sylvac probe with DRO & Square Cylinder Slip Gauges	Up to 200 mm	13µm
132	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Vernier Caliper - Plain,Digital & Dial (External,Internal & Depth) L.C.-0.01mm	Using Slip gauge & Accessories as per IS 3651- Part-I & II	0 to 1000 mm	17µm
133	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Surface Plates(Cast Iron/Granite)- Flatness	Using Digital level based on IS 12937-1990, IS7327-1991, IS2285-1991	up to 4000 X 4000 mm	2.5√L+W/125µm, where L & W in mm
134	MECHANICAL-MOBILE FORCE MEASURING SYSTEM	Pull Gauge/Force Gauge in pull mode	Using Slotted mass With hanger as per VDI/VDE 2624 Part 2.1	50 N to 500 N	1.5N
135	MECHANICAL-PRESSURE INDICATING DEVICES	Hydraulic Pressure Dial and Digital Pressure Gauges,Pressure Transmitters/Transducers Pressure Switches	Digital pressure Calibrator as per DKD-R6-1 comparison	>400 bar to 700 bar	0.6bar



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

26 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
136	MECHANICAL-PRESSURE INDICATING DEVICES	Hydraulic Pressure Dial and Digital Pressure Gauges, Pressure Transmitters/Transducers Pressure Switches	Digital pressure Calibrator as per DKD-R6-1 comparison	30 bar to 400 bar	0.15bar
137	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Pressure Dial and Digital Pressure Gauges, Pressure Transmitters/Transducers Pressure Switches	Digital pressure Calibrator -as per DKD-R6-1 comparison	>0.2 bar to 1 bar	0.005bar
138	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Pressure Dial and Digital Pressure Gauges, Pressure Transmitters/Transducers Pressure Switches	Digital pressure Calibrator as per DKD-R6-1 comparison	>1 bar to 30 bar	0.01bar
139	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Pressure Dial and Digital Pressure Gauges, Magnehelic gauges, Pressure Transmitter/Transducers Pressure Switch	Digital pressure Calibrator -as per DKD-R6-1 comparison	>20 mbar to 200 mbar	0.13mbar



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

27 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
140	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Pressure Digital and Dial Pressure Gauges , Magnehelic gauges, Pressure Transducers/Transmitters Pressure Switch	Digital low pressure Calibrator -as per DKD-R6-1 comparison	0 to 20 mbar	0.01mbar
141	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Vacuum Dial and Digital Vacuum Gauges, Vacuum Transmitters/Transducers, Vacuum Switches	Digital Vacuum Calibrator -as per DKD-R6-1 comparison	-0.87 bar to 0	0.0016bar
142	MECHANICAL-TORQUE GENERATING DEVICES	Torque Wrenches Type I Class A,B,C,D,E & Type II Class A,B,C,D,E,F,G	Using Torque Wrench calibrator as per IS/ISO 16906 :2018	2 Nm to 20 Nm	2.05%rdg
143	MECHANICAL-TORQUE GENERATING DEVICES	Torque Wrenches Type I Class A,B,C,D,E & Type II Class A,B,C,D,E,F,G	Using Torque Wrench calibrator as per IS/ISO 16906 :2018	20 Nm to 200 Nm	1.58%rdg
144	MECHANICAL-TORQUE GENERATING DEVICES	Torque Wrenches Type I Class A,B,C,D,E & Type II Class A,B,C,D,E,F,G	Using Torque Wrench calibrator as per IS/ISO 16906 :2018	200 Nm to 2000 Nm	1.75%rdg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

28 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
145	MECHANICAL-VOLUME	Volume Glass Burette	Using Weighing balance with d : 0.01 mg / 0.1 mg and distilled water based on Gravimetric method as per IS/ISO 4787 & ISO/TR 20461	10 ml to 50 ml	0.507ml
146	MECHANICAL-VOLUME	Volume Glass Pipettes (Graduated/ non graduated)	Using Weighing balance with d : 0.01 mg / 0.1 mg and distilled water based on Gravimetric method as per IS/ISO 4787 & ISO/TR 20461	50 ml to 100 ml	0.507ml
147	MECHANICAL-VOLUME	Volume Measuring cylinder /volumetric flask /Conical flask/ Beaker-Single marking & graduated	Using Weighing balance with d : 0.01 mg and distilled water based on Gravimetric method as per IS/ISO 4787 & ISO/TR 20461	1 ml to 10 ml	0.192ml
148	MECHANICAL-VOLUME	Volume Measuring cylinder /volumetric flask /Conical flask/ Beaker-Single marking & graduated	Using Weighing balance with d : 0.01 mg / 0.1 mg / 1 mg and distilled water based on Gravimetric method as per IS/ISO 4787 & ISO/TR 20461	100 ml to 500 ml	2.2ml



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

29 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
149	MECHANICAL-VOLUME	Volume Measuring cylinder /volumetric flask /Conical flask/ Beaker-Single marking & graduated	Using Weighing balance with d : 1 mg and distilled water based on Gravimetric method as per IS/ISO 4787 & ISO/TR 20461	500 ml to 1000 ml	2.6ml
150	MECHANICAL-VOLUME	Volume Micro pipettes -Piston operated	Using Weighing balance with d : 0.01 mg and distilled water based on Gravimetric method as per ISO 8655 - 6 & ISO/TR 20461	10 µl to 100 µl	0.5µl
151	MECHANICAL-VOLUME	Volume Micro pipettes -Piston operated	Using Weighing balance with d : 0.01 mg and distilled water based on Gravimetric method as per ISO 8655 - 6 & ISO/TR 20461	100 µl to 1000 µl	2.54µl
152	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class I & Coarser d =0.01 mg	Using standard weights E1&E2 class weights as per OIML R-76-1	1 g to 82 g	0.03mg
153	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class I & Coarser d =0.1 mg	Using standard weights E1&E2 class weights as per OIML R-76-1	82 g to 220 g	0.22mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO- G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

30 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
154	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class II & Coarser d =1 mg	Using standard F1 class weights as per OIML R-76-1	220 g to 1.02 kg	3.0mg
155	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class II & Coarser d =10 mg	Using standard F1 class weights as per OIML R-76-1	1.02 kg to 2.2 kg	30mg
156	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class II & Coarser d =100 mg	Using standard F1 class weights as per OIML R-76-1	2.2 kg to 32.2 kg	250mg
157	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class III & Coarser d =1 g	Using standard F1 class weights as per OIML R-76-1	32.2 kg to 200 kg	35g
158	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	1 g	0.025mg
159	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	1 mg	0.009mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

31 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
160	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	10 g	0.02mg
161	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	10 mg	0.0092mg
162	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.1 mg by ABBA Method as per OIML R-111	100 g	0.15mg
163	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	100 mg	0.0096mg
164	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	2 g	0.025mg
165	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	2 mg	0.009mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

32 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
166	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	20 g	0.03mg
167	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	20 mg	0.009mg
168	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.1 mg by ABBA Method as per OIML R-111	200 g	0.131mg
169	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	200 mg	0.025mg
170	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	5 g	0.025mg
171	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	5 mg	0.009mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	CC-2412	Page No	33 of 53
Validity	10/09/2020 to 09/09/2022	Last Amended on	-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
172	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	50 g	0.03mg
173	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	50 mg	0.01mg
174	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F1 Class Weights and coarser	Using E1 & E2 class standard weights & Balance of d : 0.01 mg by ABBA Method as per OIML R-111	500 mg	0.025mg
175	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F2 Class Weights and coarser	Using F1 class standard weights & Balance of d : 1 mg by ABBA Method as per OIML R-111	1 kg	5.0mg
176	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F2 Class Weights and coarser	Using F1 class standard weights & Balance of d : 10 mg by ABBA Method as per OIML R-111	2 kg	10.0mg
177	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of F2 Class Weights and coarser	Using F1 class standard weights & Balance of d : 1 mg by ABBA Method as per OIML R-111	500 g	3.0mg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

34 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
178	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of M1 Class Weights and coarser	Using F1 class standard weights & Balance of d : 100 mg by ABBA Method as per OIML R-111	10 kg	143.0mg
179	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of M1 Class Weights and coarser	Using F1 class standard weights & Balance of d : 100 mg by ABBA Method as per OIML R-111	20 kg	100.0mg
180	MECHANICAL-WEIGHTS	Mass /Weights- For Calibration of M1 Class Weights and coarser	Using F1 class standard weights & Balance of d : 100 mg by ABBA Method as per OIML R-111	5 kg	100.0mg
181	THERMAL-SPECIFIC HEAT & HUMIDITY	Humidity Sensor/ Transducer/Transmitter with Indicator/Digital thermo hygrometer(@25°C)	Using Temperature & Humidity meter (source Humidity Generator/Calibrator)with probe/ comparison Method	20 %RH to 95 %RH	3%RH
182	THERMAL-TEMPERATURE	Liquid in Glass Thermometer/ Dial thermometer	Using RTD (Pt-100) , 6½ DMM & Liquid bath /Direct/ comparison Method	-35 °C to 250 °C	0.75 °C
183	THERMAL-TEMPERATURE	RTD/Thermocouple / Sensor with or without temperature Indicator/ Data logger / Recorder	Using RTD (Pt-100) , 6½ DMM & Liquid bath/Dry block insert /Direct/ comparison Method	140 °C to 600 °C	0.95°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

35 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
184	THERMAL-TEMPERATURE	RTD/Thermocouple / Sensor with or without temperature Indicator/ Data logger / Recorder	Using RTD (Pt-100) , 6½ DMM & Liquid bath/Dry block insert /Direct/ comparison Method	-35 °C to 140 °C	0.45°C
185	THERMAL-TEMPERATURE	Thermocouple / Sensor with or without temperature Indicator/ Data logger / Recorder	Using R-Type Thermocouple,unive rsal calibrator & Liquid bath/Dry block insert /Direct/ comparison Method	600 °C to 1200 °C	2.15°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

36 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	1&3 Phase Energy (240V, 5A & UPF)@ 50Hz	Using Accucheck and energy source by Direct/ comparison Method	1.2 Wh to 3.6 kWh	1.18%
2	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (10Hz-1kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	100 µA to 1 A	0.55 % to 0.20 %
3	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (1kHz-5kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	1 A to 3 A	0.20 % to 0.25 %
4	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current @ (1kHz-5kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	100 µA to 1 A	0.55 % to 0.20 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

37 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC High Current @ (50Hz)	Using C.T.+6 ½ DMM FLUKE & current source by Direct/ comparison Method	20 A to 2000 A	2.10 % to 2.40 %
6	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC High Voltage @ (50Hz)	Using HV Divider & HV source by Direct/ comparison Method	1 kV to 20 kV	2.40 % to 2.60 %
7	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC High Voltage @ (50Hz)	Using HV Divider by Direct Method	20 kV to 100 kV	2.60 % to 4.05 %
8	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage @ (10Hz-20kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 mV to 1000 V	0.90 % to 0.15 %
9	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage @ (20kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 mV to 700 V	0.75 % to 0.25 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

38 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
10	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Capacitance @ (1kHz)	Using 6 ½ DMM FLUKE & capacitance box by Direct/ Comparison Method	10 mF to 100 mF	1.85 % to 4.90 %
11	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Capacitance @ (1kHz)	Using 6 ½ DMM FLUKE & capacitance box by Direct/ Comparison Method	1nF to 10 mF	5.47 % to 1.85 %
12	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current @ (10Hz-1kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	30 µA to 3 A	0.62 % to 0.09 %
13	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current @ (1kHz-5kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	30 µA to 330 mA	0.95 % to 0.2 %
14	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current @ (1kHz-5kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 mA to 10 A	0.2 % to 3.5 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

39 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
15	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current @ (45Hz-1kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	3 A to 20 A	0.09 % to 0.21 %
16	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current @ (50Hz-60Hz)	Using 5522A/9100 Calibrator(Fluke) with CC by Direct Method	10 A to 1000 A	0.5 % to 0.55 %
17	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Power 50Hz-60Hz/ 0.2pF to 1pF(Lead & Lag)/ 30V to 1000V/ 10mA to 20A	Using 5522A calibrator(Fluke)/3050 calibrator by Direct method	60 mW to 20 kW	1.8 % to 0.38 %
18	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage @ (10Hz-45Hz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	1 mV to 33 V	0.9 % to 0.05 %
19	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage @ (10kHz-100kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	30 mV to 330 mV	0.5 % to 0.15 %
20	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage @ (10kHz-100kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 mV to 330 V	0.15 % to 0.32 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

40 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
21	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage @ (45Hz-10kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	1 mV to 330 mV	0.8 % to 0.02 %
22	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage @ (45Hz-10kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 mV to 1000 V	0.02 % to 0.04 %
23	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance @ (1kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	220 pF to 330 nF	5.85 % to 0.45 %
24	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance @ (1kHz)	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 nF to 30 mF	0.45 % to 1.5 %
25	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Inductance @ (1kHz)	Using Inductance box by Direct Method	1 mH to 10 H	3%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

41 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
26	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source,Measure)	AC Current @ (10Hz-1kHz)	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	1 A to 10 A	0.20 % to 0.25 %
27	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using Shunt with 6 ½ DMM FLUKE & current source by Direct/ Comparison Method	10 A to 100 A	0.9 % to 0.53 %
28	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	100 µA to 100 mA	0.1 % to 0.07 %
29	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	100 mA to 10 A	0.07 % to 0.2 %
30	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC High Voltage	Using HV Divider with indicator &HV source by Direct/ comparison Method	1 kV to 20 kV	1.97 % to 2.10 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

42 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
31	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC High Voltage	Using HV Divider by Direct Method	20 kV to 100 kV	2.30 % to 3.06 %
32	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	1 mV to 1 V	0.42 % to 0.085 %
33	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	1 V to 1000 V	0.085 % to 0.006 %
34	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	Using 6 ½ DMM FLUKE & resistance box/ MFC by Direct/ Comparison Method	1 ohm to 100 ohm	0.15 % to 0.07 %
35	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance	Using 6 ½ DMM FLUKE, & resistance box / MFC by Direct/ Comparison Method	100 ohm to 1 Gohm	0.07 % to 2.6 %
36	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using 5522A/9100 Calibrator(Fluke) by Direct Method	10 µA to 330 mA	0.25 % to 0.02 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

43 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
37	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using 5522A/9100 Calibrator(Fluke) by Direct Method	10 A to 20 A	0.07 % to 0.026 %
38	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using 5522A/9100 Calibrator(Fluke) with CC by Direct Method	20 A to 1000 A	0.45 % to 0.64 %
39	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 mA to 10 A	0.02 % to 0.07 %
40	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using 5522A/9100 Calibrator(Fluke) by Direct Method	1 mV to 33 V	0.15 % to 0.002 %
41	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using 5522A/9100 Calibrator(Fluke) by Direct Method	33 V to 1000 V	0.002 % to 0.003 %
42	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	Using 5522A/9100 Calibrator(Fluke) by Direct Method	1 Mohm to 10 Mohm	0.004 % to 0.03 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

44 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
43	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	Using Standard Resistance Box by Direct Method	1 Gohm to 200 Gohm	1.74 % to 3.5 %
44	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	Using 5522A/9100 Calibrator(Fluke) by Direct Method	1 ohm to 1 Mohm	0.1 % to 0.004 %
45	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	Using 5522A/9100 Calibrator(Fluke) by Direct Method	10 Mohm to 330 Mohm	0.03 % to 0.35 %
46	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance	Using 5522A/9100 Calibrator(Fluke) by Direct Method	330 Mohm to 1 Gohm	0.35 % to 0.2 %
47	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscopes/ Amplitude	Using 5522A (Fluke)/3050 Calibrator by Direct Method	1 mV to 130 V	4.9 % to 0.35 %
48	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscopes/Bandwidth	Using 5522A (Fluke)/3050 Calibrator by Direct Method	50 kHz to 1.1 GHz	2.5%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

45 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
49	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscopes/Time	Using 5522A (Fluke)/3050 Calibrator by Direct Method	2 ns to 5 s	0.03 % to 0.6 %
50	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	B-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	450 °C to 1820 °C	0.8°C
51	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	E-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 1000 °C	0.15°C
52	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	J-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 1000 °C	0.6°C
53	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	K-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 1200 °C	0.6°C
54	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	N-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 1300 °C	0.6°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

46 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
55	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	R&S-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	100 °C to 1700 °C	0.65°C
56	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	RTD	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 800 °C	0.085°C
57	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	T-Type Thermocouple	Using 5522A/9100 Calibrator(Fluke) by Direct Method	-200 °C to 400 °C	0.21°C
58	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using Function Generator & MFC by Direct/ comparison Method	1 MHz to 10 MHz	0.006 % to 0.06 %
59	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using 6 ½ DMM FLUKE & MFC by Direct/ Comparison Method	10 Hz to 1000 kHz	0.06%
60	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Timer/Stop Watch/Hour meter/Time totalizer	Using time totalizer by comparison Method	2 s to 30 min	0.35 s to 1 s



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

47 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
61	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Timer/Stop Watch/Hour meter/Time totalizer	Using time totalizer by comparison Method	30 min to 24 hr	1 s to 56 s
62	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Frequency	Using 5522A/9100 Calibrator (Fluke)/Function Generator by Direct Method	10 Hz to 10 MHz	0.001 % to 0.15 %
63	MECHANICAL-ACCELERATION AND SPEED	Speed -RPM Tachometer (Contact)	Using Tachometer (source RPM generator/Calibrator)by comparison Method	>1000 RPM to 4000 RPM	10RPM
64	MECHANICAL-ACCELERATION AND SPEED	Speed -RPM Tachometer (Contact)	Using Tachometer (source RPM generator/Calibrator)by comparison Method	100 RPM to 1000 RPM	3.5RPM
65	MECHANICAL-ACCELERATION AND SPEED	Speed -RPM Tachometer (Non-Contact) / Stirrer	Using Tachometer (source RPM generator/Calibrator)by comparison Method	>4000 RPM to 47000 RPM	18RPM
66	MECHANICAL-ACCELERATION AND SPEED	Speed -RPM Tachometer (Non-Contact) / Stirrer	Using Tachometer (source RPM generator/Calibrator)by comparison Method	100 RPM to 4000 RPM	3.7RPM



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

48 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
67	MECHANICAL-ACOUSTICS	Sound Level Meter(1kHz)	Sound level generator calibrator Direct method	94dB and to 114 dB	1.8dB
68	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Height Gauge/Micro Height Gauge (Vernier/Dial/Digital) L.C.-0.1 μm	Using Slip gauge	0 to 1000 mm	10μm
69	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	V-Block-Symmetry, Parallelism & Squareness	Using Surface Plate, Test Mandrels & sylvac probe with DRO & Square Cylinder Slip Gauges	Up to 200 mm	13μm
70	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Profile projector L.C.: -0.001 mm Linear measurement	Using Glass scale by comparison method	0 to 300 mm	5μm
71	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Profile projector L.C.: -1 sec Angular measurement	Using Angular graticule scale by comparison method	0 to 360 °	1.2 min
72	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Profile projector Magnification	Using Glass scale by comparison method	10 X to 100 X	1%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

49 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
73	MECHANICAL-DIMENSION (PRECISION INSTRUMENTS)	Surface Plates(Cast Iron/Granite)- Flatness	Using Digital level based on IS 12937-1990, IS7327-1991, IS2285-1991	up to 4000 X 4000 mm	2.5vL+W/125µm, where L & W in mm
74	MECHANICAL-PRESSURE INDICATING DEVICES	Hydraulic Pressure Dial and Digital Pressure Gauges,Pressure Transmitters/Transducers Pressure Switches	Digital pressure Calibrator as per DKD-R6-1 comparison	>400 bar to 700 bar	0.6bar
75	MECHANICAL-PRESSURE INDICATING DEVICES	Hydraulic Pressure Dial and Digital Pressure Gauges, Pressure Transmitters/Transducers Pressure Switches	Digital pressure Calibrator as per DKD-R6-1 comparison	30 bar to 400 bar	0.15bar
76	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Pressure Dial and Digital Pressure Gauges, Pressure Transmitters/Transducers Pressure Switches	Digital pressure Calibrator -as per DKD-R6-1 comparison	>0.2 bar to 1 bar	0.005bar



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

50 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
77	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Pressure Dial and Digital Pressure Gauges, Pressure Transmitters/Transducers Pressure Switches	Digital pressure Calibrator as per DKD-R6-1 comparison	>1 bar to 30 bar	0.01bar
78	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Pressure Dial and Digital Pressure Gauges, Magnehelic gauges, Pressure Transmitter/Transducers Pressure Switch	Digital pressure Calibrator -as per DKD-R6-1 comparison	>20 mbar to 200 mbar	0.13mbar
79	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Pressure Digital and Dial Pressure Gauges , Magnehelic gauges, Pressure Transducers/Transmitters Pressure Switch	Digital low pressure Calibrator -as per DKD-R6-1 comparison	0 to 20 mbar	0.01mbar
80	MECHANICAL-PRESSURE INDICATING DEVICES	Pneumatic Vacuum Dial and Digital Vacuum Gauges, Vacuum Transmitters/Transducers, Vacuum Switches	Digital Vacuum Calibrator -as per DKD-R6-1 comparison	-0.87 bar to 0	0.0016bar



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

51 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
81	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class I & Coarser d =0.1 mg	Using standard weights E1&E2 class weights as per OIML R-76-1	82 g to 220 g	0.22mg
82	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class II & Coarser d =1 mg	Using standard F1 class weights as per OIML R-76-1	220 g to 1.02 kg	3.0mg
83	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class II & Coarser d =10 mg	Using standard F1 class weights as per OIML R-76-1	1.02 kg to 2.2 kg	30mg
84	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class II & Coarser d =100 mg	Using standard F1 class weights as per OIML R-76-1	2.2 kg to 32.2 kg	250mg
85	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic weighing balance Class III & Coarser d =1 g	Using standard F1 class weights as per OIML R-76-1	32.2 kg to 200 kg	35g
86	THERMAL-SPECIFIC HEAT & HUMIDITY	Indicator of Relative Humidity at a specified Single Position in environmental, Climatic ,Humidity Chamber. (@25 °C)	Using Temperature & Humidity meter (source Humidity Generator/Calibrator) with probe/ comparison Method	20 %RH to 95 %RH	3%RH



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,, CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

52 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
87	THERMAL-TEMPERATURE	RTD/Thermocouple / Sensor with or without temperature Indicator/ Data logger / Recorder	Using RTD (Pt-100) , 6½ DMM & Liquid bath/Dry block insert /Direct/ comparison Method	140 °C to 600 °C	0.95°C
88	THERMAL-TEMPERATURE	RTD/Thermocouple / Sensor with or without temperature Indicator/ Data logger / Recorder	Using RTD (Pt-100) , 6½ DMM & Liquid bath/Dry block insert /Direct/ comparison Method	-35 °C to 140 °C	0.45°C
89	THERMAL-TEMPERATURE	Spatial mapping using multi sensor & data logger of Freezer, cold chamber Oven, Furnace,Coating oven	Using PT-100 Sensor & Thermocouple & Data logger/ comparison Method	-40 °C to 400 °C	2.60°C
90	THERMAL-TEMPERATURE	Spatial mapping using multi sensor & data logger of Oven, Furnace, Coating oven	Using Thermocouple & Data logger/ comparison Method	400 °C to 1200 °C	3.80°C
91	THERMAL-TEMPERATURE	Temperature Indicator (at single specified Position) of Freezer, cold chamber, Oven, Furnace	Using RTD (Pt-100) & 6½ DMM /Direct/ comparison Method	-65 °C to 600 °C	1.60°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

UNIVERSAL CALIBRATION SERVICES PVT. LIMITED, PLOT NO-
G-43/1&2,G-44/1&2,G-BLOCK, AJANTA NAGAR,BEHIND KASTURI MARKET,MIDC,,
CHINCHWAD, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2412

Page No

53 of 53

Validity

10/09/2020 to 09/09/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
92	THERMAL-TEMPERATURE	Temperature Indicator (at single specified Position) of Oven, Furnace	Using Thermocouple & universal calibrator/ comparison Method	600 °C to 1200 °C	2.92°C
93	THERMAL-TEMPERATURE	Thermocouple / Sensor with or without temperature Indicator/ Data logger / Recorder	Using R-Type Thermocouple,unive rsal calibrator & Liquid bath/Dry block insert /Direct/ comparison Method	600 °C to 1200 °C	2.15°C

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.